Introduction

The Tewitro® TW 24 is the only device to measure the water evaporation from **cultured tissue sets** (wells in a plate with medium) in up to 24 wells simultaneously with the worldwide most used **open chamber** measurement of the **Tewameter**®.

The Measuring Principle

The Tewitro® TW 24 can be used in up to **24 wells**plates (6x4). Each measurement inlet features two
sensor pairs constantly measuring temperature and
relative humidity, thus measuring in an indirect way
the gradient of the water evaporation from the surface
of cell cultures at the bottom of the well. This gradient
equals the Transepidermal Water Loss typically measured on the in vivo skin surface in g/h/m².

Fields of Application

For each product to be applied to the skin, **safety measurements** are indispensable. The use of cultured cells sets for long-term safety tests is a quick and easy method to **avoid using animals or volunteers**.
•


- Only way to study skin permeability and dermal absorption necessary for safety and efficacy • testing.
- Well established in different guidelines around the globe and approved by institutions such as ECVAM (European Centre for the Validation of Alternatives to Animal Testing).

Whenever the barrier is damaged, the evaporation rate will increase immediately. Also for claims related to **barrier improvement/enhancing**, tests on cultured cells are needed. Products meant to keep the barrier intact or to even improve it, can be identified by a stable TEWL after their application.

Advantages

- All values and results of each sensor are clearly represented in the convenient software.
- The probe can measure continuously over longer periods.
- The software indicates that the TEWL has reached a stable threshold value.
- You can be sure to apply the tested products in repeated tests to cell cultures under the same conditions.
- Control value of all wells used is recorded before application, so that the effects measured afterwards can be surely attributed to the product.
- Extremely time- and manpower-saving.
- Suitable for Epiderm[®], Episkin[®]. Please ask for other adaptations.
- Tewitro® TW 24R for TEWL measurements on full thickness reconstructed epidermis (e.g. Phenion®FT by Henkel).
- Available for C+K MPA-systems (with Tewitro® software).









Technical Data

Dimensions: 113 (W) x 170 (L) x 32 (H) mm, weight: 300 g, cable length: approx. 1.20 m, frame material: anodized aluminum (AlMg 3), power consumption: max. 12 V; In full equipment: 24 sensor pairs (48 single sensors)

Measurement range: Temperature: $0 - 50^{\circ}$ C, resolution: typ. 0.015° C, RH: 0% - 100% RH, resolution: typ. 0.01% RH, TEWL: 0 to 320 g/h/m²; Measurement uncertainty: for $20-50^{\circ}$ C and RH $\leq 80\%$: typ. $\pm 1.5\%$ RH, max. $\pm 2\%$ RH, typ. $\pm 0.1^{\circ}$ C, max. $\pm 0.3^{\circ}$ C; Operating conditions: T: $5-40^{\circ}$ C RH: 30-70% Technical changes may be made without prior notice.

Courage+Khazaka electronic GmbH since 1986 Mathias-Brüggen-Str. 91 · 50829 Köln · GERMANY

phone +49 221 95 64 99 0 · fax +49 221 95 64 99 1 info@courage-khazaka.de · www.courage-khazaka.de

